

Position the open

end of the Eyelet

Hook toward the

front of the shoe.









Place the SaeboStep around the ankle with the Front Dial centered to your shoe and the blue Saebo logo positioned at the top.

SECURE ANKLE CUFF



Grab Pull Tab that is attached to the end of the Ankle Cuff Cord and connect to magnetic base. You will hear a click when secured.

Turn Ankle Cuff Dial clockwise until the Cuff is fit snugly.





Wind the dial by turning clockwise until the Front Cord is tight and you have the lift and angle desired.







1 Rotate Ankle Cuff so Magnetic Fastener and Ankle Cuff Dial are near the front for easier access.

2 Attach Magnetic Fastener to Base.

3 Tighten Ankle Cuff Cord by winding the Ankle Cuff Dial clockwise.



4 Rotate brace back to the proper position with the Saebo logo facing the front.

5 Finish tightening the Ankle Cuff Cord by winding the Ankle Cuff Dial clockwise until you feel a snug fit.

SEE THE **VIDEOS**

Visit our website to see instructional videos including how to punch new eyelets into practically any type of footwear.

www.saebo.com/saebostep

Consult your doctor concerning your condition and the use of this brace.



European Healthcare & Device Solutions (Ireland) Ltd. info@europeandevicesolutions.co.uk www.europeandevicesolutions.co.uk

Australian Sponsor: Ausmedic Australia Pty Ltd Unit 3, 3 Basalt Road Pemulwuy NSW 2145 1300 473 422 Performhealth.com.au

SECURE ANKLE CUFF CONTINUED



Note 1: Make sure the Ankle Cuff Cord is not twisted.



Note 2: The Cuff on the dial side should overlap the Cuff on the magnetic side.



Saebo, Inc. 2459 Wilkinson Blvd., Suite 120-B Charlotte, NC 28208

888-942-5751 sales@saebo.com www.saebo.com

EC REP

